

## Sample Statistical Summary

MAPEP-11-GrW25

Analyte	T(1)	A(2)	Grand(3) Mean	Std Dev	Ref Value	Ref Unc	Acceptance Range
Radiological							Units: (Bq/L)
Gross alpha	56	50	0.899	0.205	0.866	0.009	0.260 - 1.472
Gross beta	57	55	4.87	0.78	4.81	0.08	2.41 - 7.22

**Note:** (1) T = Total number of laboratories reporting analyte.  
(2) A = Number of laboratories with 'Acceptable' performance.  
(3) Mean excludes values derived as total metals and values indicated as "Not Acceptable".

### Gross Alpha Flags:

A = Result acceptable, Bias 2, i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).

N = Result not acceptable, Bias > +/- 70% or the reported result is not statistically positive at two standard deviations (Result/Uncertainty)

### Gross Beta Flags:

A = Result acceptable, Bias 2, i.e., the range encompassing the result, plus or minus the total uncertainty at two standard deviations, does not include zero).

N = Result not acceptable, Bias > +/- 50% or the reported result is not statistically positive at two standard deviations (Result/Uncertainty)